

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026959.D
 Acq On : 11 Feb 2022 11:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 12 04:26:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.544	0.527	3.1	90	0.00
3 P	Chloromethane	0.506	0.483	4.5	91	0.00
4 C	Vinyl Chloride	0.529	0.510	3.6#	90	0.00
5 T	Bromomethane	0.225	0.215	4.4	96	0.00
6 T	Chloroethane	0.281	0.297	-5.7	90	0.00
7 T	Trichlorofluoromethane	0.779	0.748	4.0	89	0.00
8 T	Diethyl Ether	0.289	0.287	0.7	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.511	0.500	2.2	92	0.00
10 T	Methyl Iodide	0.656	0.635	3.2	87	0.00
11 T	Tert butyl alcohol	0.108	0.088	18.5	79	-0.03
12 CM	1,1-Dichloroethene	0.512	0.473	7.6#	87	0.00
13 T	Acrolein	0.100	0.117	-17.0	110	0.00
14 T	Allyl chloride	0.906	0.892	1.5	92	0.00
15 T	Acrylonitrile	0.311	0.306	1.6	92	0.00
16 T	Acetone	0.264	0.314	-18.9	117	0.00
17 T	Carbon Disulfide	1.449	1.295	10.6	85	0.00
18 T	Methyl Acetate	0.758	0.720	5.0	93	0.00
19 T	Methyl tert-butyl Ether	1.811	1.773	2.1	93	0.00
20 T	Methylene Chloride	0.571	0.551	3.5	95	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.521	3.2	90	0.00
22 T	Diisopropyl ether	1.825	1.812	0.7	93	0.00
23 T	Vinyl Acetate	1.534	1.585	-3.3	94	0.00
24 P	1,1-Dichloroethane	0.987	0.973	1.4	93	0.00
25 T	2-Butanone	0.425	0.441	-3.8	97	0.00
26 T	2,2-Dichloropropane	0.667	0.732	-9.7	100	0.00
27 T	cis-1,2-Dichloroethene	0.618	0.619	-0.2	93	0.00
28 T	Bromochloromethane	0.409	0.422	-3.2	97	0.00
29 T	Tetrahydrofuran	0.289	0.277	4.2	91	0.00
30 C	Chloroform	1.026	1.004	2.1#	94	0.00
31 T	Cyclohexane	0.954	0.905	5.1	89	0.00
32 T	1,1,1-Trichloroethane	0.905	0.868	4.1	89	0.00
33 S	1,2-Dichloroethane-d4	0.600	0.573	4.5	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.298	0.284	4.7	92	0.00
36 T	1,1-Dichloropropene	0.470	0.469	0.2	92	0.00
37 T	Ethyl Acetate	0.534	0.538	-0.7	93	0.00
38 T	Carbon Tetrachloride	0.493	0.487	1.2	90	0.00
39 T	Methylcyclohexane	0.576	0.590	-2.4	92	0.00
40 TM	Benzene	1.370	1.370	0.0	92	0.00
41 T	Methacrylonitrile	0.289	0.295	-2.1	92	0.00
42 TM	1,2-Dichloroethane	0.483	0.512	-6.0	96	0.00
43 T	Isopropyl Acetate	0.822	0.851	-3.5	92	0.00
44 TM	Trichloroethene	0.411	0.402	2.2	89	0.00
45 C	1,2-Dichloropropane	0.347	0.355	-2.3#	94	0.00
46 T	Dibromomethane	0.240	0.252	-5.0	96	0.00
47 T	Bromodichloromethane	0.459	0.491	-7.0	95	0.00
48 T	Methyl methacrylate	0.407	0.420	-3.2	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026959.D
 Acq On : 11 Feb 2022 11:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 12 04:26:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	79	0.00
50 S	Toluene-d8	1.093	1.046	4.3	92	0.00
51 T	4-Methyl-2-Pentanone	0.498	0.506	-1.6	93	0.00
52 CM	Toluene	0.867	0.858	1.0#	92	0.00
53 T	t-1,3-Dichloropropene	0.446	0.516	-15.7	99	0.00
54 T	cis-1,3-Dichloropropene	0.507	0.577	-13.8	98	0.00
55 T	1,1,2-Trichloroethane	0.328	0.346	-5.5	95	0.00
56 T	Ethyl methacrylate	0.520	0.553	-6.3	95	0.00
57 T	1,3-Dichloropropane	0.564	0.602	-6.7	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.267	0.263	1.5	83	0.00
59 T	2-Hexanone	0.364	0.374	-2.7	93	0.00
60 T	Dibromochloromethane	0.338	0.367	-8.6	94	0.00
61 T	1,2-Dibromoethane	0.347	0.370	-6.6	96	0.00
62 S	4-Bromofluorobenzene	0.385	0.365	5.2	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.502	0.480	4.4	86	0.00
65 PM	Chlorobenzene	1.034	1.027	0.7	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.392	-3.2	95	0.00
67 C	Ethyl Benzene	1.830	1.858	-1.5#	93	0.00
68 T	m/p-Xylenes	0.704	0.702	0.3	93	0.00
69 T	o-Xylene	0.690	0.683	1.0	92	0.00
70 T	Styrene	1.119	1.138	-1.7	93	0.00
71 P	Bromoform	0.263	0.281	-6.8	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.938	3.996	-1.5	92	0.00
74 T	N-amyl acetate	1.574	1.614	-2.5	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.069	1.083	-1.3	97	0.00
76 T	1,2,3-Trichloropropane	1.095	1.129	-3.1	95	0.00
77 T	Bromobenzene	0.936	0.944	-0.9	93	0.00
78 T	n-propylbenzene	4.366	4.530	-3.8	93	0.00
79 T	2-Chlorotoluene	2.698	2.698	0.0	94	0.00
80 T	1,3,5-Trimethylbenzene	3.241	3.273	-1.0	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.317	0.354	-11.7	97	0.00
82 T	4-Chlorotoluene	3.026	3.061	-1.2	92	0.00
83 T	tert-Butylbenzene	3.089	3.132	-1.4	94	0.00
84 T	1,2,4-Trimethylbenzene	3.269	3.273	-0.1	93	0.00
85 T	sec-Butylbenzene	3.916	4.039	-3.1	94	0.00
86 T	p-Isopropyltoluene	3.318	3.439	-3.6	94	0.00
87 T	1,3-Dichlorobenzene	1.715	1.737	-1.3	94	0.00
88 T	1,4-Dichlorobenzene	1.730	1.739	-0.5	94	0.00
89 T	n-Butylbenzene	2.714	2.889	-6.4	96	0.00
90 T	Hexachloroethane	0.479	0.531	-10.9	98	0.00
91 T	1,2-Dichlorobenzene	1.655	1.698	-2.6	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.263	0.266	-1.1	89	0.00
93 T	1,2,4-Trichlorobenzene	1.045	1.111	-6.3	96	0.00
94 T	Hexachlorobutadiene	0.442	0.452	-2.3	93	0.00
95 T	Naphthalene	3.659	3.805	-4.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
Data File : VX026959.D
Acq On : 11 Feb 2022 11:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 12 04:26:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.058	1.113	-5.2	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6