



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021122\
 Data File : VX026958.D
 Acq On : 11 Feb 2022 09:27
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Title : SW846 8260
 Last Update : Tue Feb 08 07:07:46 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	4561	PASS
75	95	30	60	50.7	12526	PASS
95	95	100	100	100.0	24704	PASS
96	95	5	9	6.5	1607	PASS
173	174	0.00	2	1.0	173	PASS
174	95	50	100	71.5	17673	PASS
175	174	5	9	7.9	1400	PASS
176	174	95	101	98.1	17344	PASS
177	176	5	9	6.2	1073	PASS