

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021123\
 Data File : VX034094.D
 Acq On : 09 Feb 2023 10:28
 Operator : JC/MD
 Sample : VX0211WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0211WBL01

Quant Time: Feb 10 01:03:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	127722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177676	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158294	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61225	49.724	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.440%
35) Dibromofluoromethane	5.385	113	56436	49.909	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.646	98	200827	49.293	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.580%
62) 4-Bromofluorobenzene	11.079	95	72454	49.793	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021123\
Data File : VX034094.D
Acq On : 09 Feb 2023 10:28
Operator : JC/MD
Sample : VX0211WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0211WBL01

Quant Time: Feb 10 01:03:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

