

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021123\
 Data File : VX034097.D
 Acq On : 09 Feb 2023 11:40
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 10 01:06:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126391	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173813	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79001	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61148	50.184	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.360%	
35) Dibromofluoromethane	5.385	113	55462	50.137	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.280%	
50) Toluene-d8	8.647	98	199962	50.171	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.340%	
62) 4-Bromofluorobenzene	11.079	95	70886	49.798	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.600%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021123\
Data File : VX034097.D
Acq On : 09 Feb 2023 11:40
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 10 01:06:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

