

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021123\
 Data File : VX034099.D
 Acq On : 09 Feb 2023 12:27
 Operator : JC/MD
 Sample : 01509-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-04B-18-28-020823

Quant Time: Feb 10 01:07:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	124856	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173375	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	154848	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75615	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60984	50.665	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.320%
35) Dibromofluoromethane	5.385	113	54819	49.681	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	8.647	98	197392	49.652	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.300%
62) 4-Bromofluorobenzene	11.079	95	70411	49.590	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.180%
Target Compounds						Qvalue
7) Trichlorofluoromethane	1.892	101	1686	1.056	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021123\
Data File : VX034099.D
Acq On : 09 Feb 2023 12:27
Operator : JC/MD
Sample : 01509-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-04B-18-28-020823

Quant Time: Feb 10 01:07:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

