

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021123\
 Data File : VX034102.D
 Acq On : 09 Feb 2023 13:43
 Operator : JC/MD
 Sample : 01462-13
 Misc : 1.20g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOT-VOC-GRAB-12

Quant Time: Feb 09 14:30:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	117448	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.757	114	161926	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	142986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75702	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	55704	49.197	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.400%
35) Dibromofluoromethane	5.385	113	47656	46.243	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.480%
50) Toluene-d8	8.647	98	181574	48.902	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.800%
62) 4-Bromofluorobenzene	11.085	95	65132	49.115	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021123\
Data File : VX034102.D
Acq On : 09 Feb 2023 13:43
Operator : JC/MD
Sample : 01462-13
Misc : 1.20g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOT-VOC-GRAB-12

Quant Time: Feb 09 14:30:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

