



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021123\
Data File : VX034091.D
Acq On : 09 Feb 2023 08:54
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Title : SW846 8260
Last Update : Fri Feb 03 23:25:45 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	5744	PASS
75	95	30	60	45.6	16049	PASS
95	95	100	100	100.0	35157	PASS
96	95	5	9	7.1	2498	PASS
173	174	0.00	2	1.3	442	PASS
174	95	50	100	96.6	33949	PASS
175	174	5	9	7.6	2585	PASS
176	174	95	101	95.1	32272	PASS
177	176	5	9	6.7	2158	PASS