

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
 Data File : VX044895.D
 Acq On : 11 Feb 2025 11:00
 Operator : JC/MD
 Sample : VX0211WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0211WBL01

Quant Time: Feb 12 02:37:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	98496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	199075	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	183423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	79632	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	77084	53.464	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.920%
35) Dibromofluoromethane	5.379	113	65973	50.971	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.940%
50) Toluene-d8	8.647	98	245936	50.249	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.500%
62) 4-Bromofluorobenzene	11.079	95	86470	52.406	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
Data File : VX044895.D
Acq On : 11 Feb 2025 11:00
Operator : JC/MD
Sample : VX0211WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0211WBL01

Quant Time: Feb 12 02:37:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

