

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
 Data File : VX044898.D
 Acq On : 11 Feb 2025 12:11
 Operator : JC/MD
 Sample : Q1289-04ME
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-DRUMS-BME

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/12/2025
 Supervised By : Mahesh Dadoda 02/12/2025

Quant Time: Feb 12 06:52:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	95409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	191599	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	170333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73984	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72846	52.159	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.320%			
35) Dibromofluoromethane	5.379	113	61567	49.423	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.840%			
50) Toluene-d8	8.647	98	235387	49.970	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.940%			
62) 4-Bromofluorobenzene	11.079	95	79898	50.313	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.620%			
Target Compounds						
16) Acetone	2.404	43	1065m	1.897	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
Data File : VX044898.D
Acq On : 11 Feb 2025 12:11
Operator : JC/MD
Sample : Q1289-04ME
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-DRUMS-BME

Manual Integrations
APPROVED

Reviewed By : John Carlone 02/12/2025
Supervised By : Mahesh Dadoda 02/12/2025

Quant Time: Feb 12 06:52:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

