

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
 Data File : VX044900.D
 Acq On : 11 Feb 2025 12:57
 Operator : JC/MD
 Sample : Q1283-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB192-HYD-20250131

Quant Time: Feb 12 02:41:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	91747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	188311	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	169014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73561	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	71542	53.270	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.540%		
35) Dibromofluoromethane	5.379	113	60372	49.310	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.620%		
50) Toluene-d8	8.647	98	229287	49.525	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.060%		
62) 4-Bromofluorobenzene	11.079	95	79624	51.016	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.040%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1018	1.886	ug/l	97
17) Carbon Disulfide	2.508	76	1018	0.314	ug/l #	76
60) Dibromochloromethane	9.525	129	1563	1.157	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
Data File : VX044900.D
Acq On : 11 Feb 2025 12:57
Operator : JC/MD
Sample : Q1283-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB192-HYD-20250131

Quant Time: Feb 12 02:41:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

