

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
 Data File : VX044903.D
 Acq On : 11 Feb 2025 14:06
 Operator : JC/MD
 Sample : Q1283-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-480-482

Quant Time: Feb 12 02:42:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	92053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	188552	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	173671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	77145	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	73292	54.392	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.780%	
35) Dibromofluoromethane	5.379	113	61697	50.328	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.660%	
50) Toluene-d8	8.647	98	234124	50.505	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.020%	
62) 4-Bromofluorobenzene	11.079	95	83740	53.584	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.160%	
Target Compounds						Qvalue
16) Acetone	2.380	43	3617	6.677	ug/l	96
17) Carbon Disulfide	2.508	76	1460	0.449	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
Data File : VX044903.D
Acq On : 11 Feb 2025 14:06
Operator : JC/MD
Sample : Q1283-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-480-482

Quant Time: Feb 12 02:42:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

