

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
 Data File : VX044912.D
 Acq On : 11 Feb 2025 17:33
 Operator : JC/MD
 Sample : Q1327-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-520-522

Quant Time: Feb 12 02:45:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	88393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	180540	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	161074	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71210	55.035	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.060%
35) Dibromofluoromethane	5.379	113	59009	50.271	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.540%
50) Toluene-d8	8.647	98	220692	49.720	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.440%
62) 4-Bromofluorobenzene	11.079	95	74852	50.023	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.040%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	7866	15.122	ug/l	99
17) Carbon Disulfide	2.502	76	6303	2.017	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021125\
Data File : VX044912.D
Acq On : 11 Feb 2025 17:33
Operator : JC/MD
Sample : Q1327-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-520-522

Quant Time: Feb 12 02:45:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

