

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021219\
 Data File : VX007493.D
 Acq On : 12 Feb 2019 17:16
 Operator : JC/SP
 Sample : K1457-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-LIQUID-02

Quant Time: Feb 13 01:13:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	543962	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	842724	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	777322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	386113	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	285226	49.40	ug/l	0.00
Spiked Amount			Recovery	=	98.80%	
35) Dibromofluoromethane	5.51	113	263831	48.12	ug/l	0.00
Spiked Amount			Recovery	=	96.24%	
50) Toluene-d8	8.71	98	996501	48.98	ug/l	0.00
Spiked Amount			Recovery	=	97.96%	
62) 4-Bromofluorobenzene	11.13	95	361718	48.70	ug/l	0.00
Spiked Amount			Recovery	=	97.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	132958	50.272	ug/l	98
25) 2-Butanone	4.70	43	44173	11.736	ug/l	99
31) Cyclohexane	5.59	56	59993	7.472	ug/l #	88
39) Methylcyclohexane	7.46	83	142088	15.532	ug/l	99
52) Toluene	8.79	92	15637	1.088	ug/l	91
59) 2-Hexanone	9.49	43	7996	1.323	ug/l	76
64) Tetrachloroethene	9.34	164	8847	1.208	ug/l	92
67) Ethyl Benzene	10.25	91	44674	1.635	ug/l	94
68) m/p-Xylenes	10.36	106	38029	3.502	ug/l	97
73) Isopropylbenzene	11.02	105	36472	1.390	ug/l	98
78) n-propylbenzene	11.36	91	167434	5.819	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	170993	7.966	ug/l	100
84) 1,2,4-Trimethylbenzene	11.80	105	413319	18.965	ug/l	98
89) n-Butylbenzene	12.38	91	76151	3.812	ug/l #	21

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021219\
Data File : VX007493.D
Acq On : 12 Feb 2019 17:16
Operator : JC/SP
Sample : K1457-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-LIQUID-02

Quant Time: Feb 13 01:13:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

