

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021220\
 Data File : VX014921.D
 Acq On : 12 Feb 2020 13:31
 Operator : JC/SP
 Sample : L1495-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW105-200211

Quant Time: Feb 13 11:26:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	437475	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	678282	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	652516	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	338667	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	249226	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
35) Dibromofluoromethane	5.49	113	207034	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
50) Toluene-d8	8.71	98	812387	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
62) 4-Bromofluorobenzene	11.13	95	301130	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	408701	82.633	ug/l	89
44) Trichloroethene	7.21	130	169598	32.227	ug/l	95
64) Tetrachloroethene	9.33	164	231800	37.995	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021220\
Data File : VX014921.D
Acq On : 12 Feb 2020 13:31
Operator : JC/SP
Sample : L1495-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW105-200211

Quant Time: Feb 13 11:26:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

