

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021220\
 Data File : VX014923.D
 Acq On : 12 Feb 2020 14:16
 Operator : JC/SP
 Sample : L1495-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW002-200211

Quant Time: Feb 13 04:00:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	439534	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	679762	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	654251	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	341222	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	251233	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
35) Dibromofluoromethane	5.49	113	211031	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
50) Toluene-d8	8.71	98	818099	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
62) 4-Bromofluorobenzene	11.13	95	303768	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
Target Compounds						
4) Vinyl Chloride	1.41	62	130786	25.726	ug/l	95
6) Chloroethane	1.73	64	7384	2.270	ug/l	99
21) trans-1,2-Dichloroethene	3.17	96	5721	1.287	ug/l	93
27) cis-1,2-Dichloroethene	4.59	96	374908	75.446	ug/l	89
44) Trichloroethene	7.21	130	9026	1.711	ug/l	82
64) Tetrachloroethene	9.33	164	6544	1.070	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021220\
Data File : VX014923.D
Acq On : 12 Feb 2020 14:16
Operator : JC/SP
Sample : L1495-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW002-200211

Quant Time: Feb 13 04:00:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

