

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021220\
 Data File : VX014924.D
 Acq On : 12 Feb 2020 14:39
 Operator : JC/SP
 Sample : L1495-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY036MW016-200211

Quant Time: Feb 13 04:02:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	423261	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	660416	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	637099	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	338065	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	243664	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
35) Dibromofluoromethane	5.49	113	205902	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
50) Toluene-d8	8.71	98	797805	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
62) 4-Bromofluorobenzene	11.13	95	301586	54.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.04%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	7117	1.454	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	14096	2.946	ug/l	85
44) Trichloroethene	7.21	130	8847	1.727	ug/l	93
64) Tetrachloroethene	9.33	164	14552	2.443	ug/l	94
83) tert-Butylbenzene	11.76	119	7559	0.451	ug/l	88
88) 1,4-Dichlorobenzene	12.09	146	12654	1.167	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021220\
Data File : VX014924.D
Acq On : 12 Feb 2020 14:39
Operator : JC/SP
Sample : L1495-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY036MW016-200211

Quant Time: Feb 13 04:02:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

