


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX021220\
 Data File : VX014915.D
 Acq On : 12 Feb 2020 10:13
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Title : SW846 8260
 Last Update : Mon Feb 10 12:09:20 2020



AutoFind: Scans 1680, 1681, 1682; Background Corrected with Scan 1674

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	12130	PASS
75	95	30	60	47.2	31752	PASS
95	95	100	100	100.0	67328	PASS
96	95	5	9	6.5	4363	PASS
173	174	0.00	2	1.3	809	PASS
174	95	50	100	89.2	60053	PASS
175	174	5	9	7.3	4374	PASS
176	174	95	101	99.3	59603	PASS
177	176	5	9	6.3	3743	PASS