

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
 Data File : VX034109.D
 Acq On : 10 Feb 2023 13:09
 Operator : JC/MD
 Sample : 01445-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-26-020323

Quant Time: Feb 10 13:32:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	119915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167164	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	148565	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59803	51.731	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.460%		
35) Dibromofluoromethane	5.385	113	53141	49.950	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.900%		
50) Toluene-d8	8.647	98	189835	49.525	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.060%		
62) 4-Bromofluorobenzene	11.079	95	67011	48.949	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 97.900%		
Target Compounds						
						Qvalue
16) Acetone	2.373	43	37427	76.653	ug/l	96
20) Methylene Chloride	2.794	84	4327	3.874	ug/l	89
25) 2-Butanone	4.562	43	5692	8.006	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	2037	1.656	ug/l	89
37) Ethyl Acetate	4.714	43	14129	10.398	ug/l	98
43) Isopropyl Acetate	6.336	43	61521	29.031	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
Data File : VX034109.D
Acq On : 10 Feb 2023 13:09
Operator : JC/MD
Sample : 01445-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-26-020323

Quant Time: Feb 10 13:32:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

