

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
 Data File : VX034114.D
 Acq On : 10 Feb 2023 15:05
 Operator : JC/MD
 Sample : 01519-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10-GW-698-700

Quant Time: Feb 11 02:50:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	117939	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	164643	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72021	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59236	52.099	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.200%
35) Dibromofluoromethane	5.385	113	51962	49.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.180%
50) Toluene-d8	8.647	98	187574	49.685	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	66365	49.219	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.440%
Target Compounds						
16) Acetone	2.380	43	1084	2.257	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
Data File : VX034114.D
Acq On : 10 Feb 2023 15:05
Operator : JC/MD
Sample : 01519-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-GW-698-700

Quant Time: Feb 11 02:50:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

