

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
 Data File : VX034115.D
 Acq On : 10 Feb 2023 15:28
 Operator : JC/MD
 Sample : 01519-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10-GW-718-720

Quant Time: Feb 11 02:50:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	118115	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	162976	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70174	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57342	50.358	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.720%	
35) Dibromofluoromethane	5.385	113	51959	50.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.180%	
50) Toluene-d8	8.647	98	184256	49.305	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.600%	
62) 4-Bromofluorobenzene	11.079	95	64889	48.617	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.240%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
Data File : VX034115.D
Acq On : 10 Feb 2023 15:28
Operator : JC/MD
Sample : 01519-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10-GW-718-720

Quant Time: Feb 11 02:50:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

