

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
 Data File : VX034117.D
 Acq On : 10 Feb 2023 16:15
 Operator : JC/MD
 Sample : 01516-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-01-160-180-020923

Quant Time: Feb 11 00:35:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	111782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	155476	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65327	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	55559	51.557	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.120%	
35) Dibromofluoromethane	5.385	113	50174	50.707	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.420%	
50) Toluene-d8	8.647	98	176208	49.426	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.860%	
62) 4-Bromofluorobenzene	11.079	95	61142	48.019	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.040%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1496	3.287	ug/l	91
44) Trichloroethene	7.129	130	1449	1.201	ug/l	97
52) Toluene	8.720	92	4620	1.816	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
Data File : VX034117.D
Acq On : 10 Feb 2023 16:15
Operator : JC/MD
Sample : 01516-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-01-160-180-020923

Quant Time: Feb 11 00:35:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

