

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
 Data File : VX034120.D
 Acq On : 10 Feb 2023 17:25
 Operator : JC/MD
 Sample : 01516-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-020923

Quant Time: Feb 11 00:36:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	111549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	155191	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139311	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69164	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56563	52.598	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.200%
35) Dibromofluoromethane	5.385	113	49256	49.870	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.740%
50) Toluene-d8	8.647	98	177543	49.892	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.780%
62) 4-Bromofluorobenzene	11.079	95	62664	49.305	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
Data File : VX034120.D
Acq On : 10 Feb 2023 17:25
Operator : JC/MD
Sample : 01516-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-020923

Quant Time: Feb 11 00:36:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

