

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
 Data File : VX034129.D
 Acq On : 10 Feb 2023 20:54
 Operator : JC/MD
 Sample : 01481-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-42-020723

Quant Time: Feb 11 00:41:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	153260	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62296	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55559	52.099	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.200%
35) Dibromofluoromethane	5.385	113	48905	50.139	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.280%
50) Toluene-d8	8.646	98	171473	48.793	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	59623	47.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.000%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	23236	51.589	ug/l	94
20) Methylene Chloride	2.794	84	7271	7.056	ug/l	93
37) Ethyl Acetate	4.708	43	12883	10.341	ug/l	97
43) Isopropyl Acetate	6.336	43	56421	29.040	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
Data File : VX034129.D
Acq On : 10 Feb 2023 20:54
Operator : JC/MD
Sample : 01481-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-42-020723

Quant Time: Feb 11 00:41:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

