

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
 Data File : VX034130.D
 Acq On : 10 Feb 2023 21:18
 Operator : JC/MD
 Sample : 01481-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-46-020723

Quant Time: Feb 11 00:42:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	114845	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	160086	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141418	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70721	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57156	51.624	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.240%	
35) Dibromofluoromethane	5.385	113	51112	50.167	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.340%	
50) Toluene-d8	8.647	98	182701	49.771	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.540%	
62) 4-Bromofluorobenzene	11.079	95	65670	50.090	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.180%	
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	23879	51.065	ug/l	94
20) Methylene Chloride	2.788	84	7391	6.909	ug/l	91
37) Ethyl Acetate	4.714	43	13775	10.586	ug/l	98
43) Isopropyl Acetate	6.336	43	58095	28.627	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
Data File : VX034130.D
Acq On : 10 Feb 2023 21:18
Operator : JC/MD
Sample : 01481-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-46-020723

Quant Time: Feb 11 00:42:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

