

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
 Data File : VX034131.D
 Acq On : 10 Feb 2023 21:41
 Operator : JC/MD
 Sample : 01481-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JPP-48-020723

Quant Time: Feb 11 00:42:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	113431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	159103	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57379	52.471	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.940%
35) Dibromofluoromethane	5.385	113	51204	50.568	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.140%
50) Toluene-d8	8.647	98	178369	48.891	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.780%
62) 4-Bromofluorobenzene	11.079	95	63713	48.897	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.800%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	23924	51.799	ug/l	99
20) Methylene Chloride	2.788	84	7597	7.190	ug/l	88
37) Ethyl Acetate	4.708	43	13754	10.635	ug/l	98
43) Isopropyl Acetate	6.336	43	57802	28.658	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021223\
Data File : VX034131.D
Acq On : 10 Feb 2023 21:41
Operator : JC/MD
Sample : 01481-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JPP-48-020723

Quant Time: Feb 11 00:42:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

