

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
 Data File : VX040187.D
 Acq On : 12 Feb 2024 10:23
 Operator : JC/MD
 Sample : VX0212WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0212WBL01

Quant Time: Feb 12 14:21:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	145968	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	140702	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55675	43.384	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.760%
35) Dibromofluoromethane	5.385	113	46517	45.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.260%
50) Toluene-d8	8.647	98	178614	46.842	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.680%
62) 4-Bromofluorobenzene	11.079	95	71692	46.100	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
Data File : VX040187.D
Acq On : 12 Feb 2024 10:23
Operator : JC/MD
Sample : VX0212WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0212WBL01

Quant Time: Feb 12 14:21:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

