

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
 Data File : VX040191.D
 Acq On : 12 Feb 2024 11:57
 Operator : JC/MD
 Sample : P1411-01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-11929

Quant Time: Feb 12 14:22:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	93876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	157686	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	147731	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83021	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60935	43.631	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.260%
35) Dibromofluoromethane	5.385	113	46337	42.079	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.160%
50) Toluene-d8	8.647	98	191821	46.567	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.140%
62) 4-Bromofluorobenzene	11.085	95	83721	49.834	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.660%
Target Compounds						
					Qvalue	
18) Methyl Acetate	2.727	43	3690	2.400	ug/l	98
39) Methylcyclohexane	7.372	83	6299	3.514	ug/l	97
67) Ethyl Benzene	10.195	91	1589	0.289	ug/l	97
73) Isopropylbenzene	10.963	105	18367	3.141	ug/l	100
78) n-propylbenzene	11.305	91	40677	5.746	ug/l	98
85) sec-Butylbenzene	11.890	105	35278	5.688	ug/l #	71
89) n-Butylbenzene	12.335	91	78006	15.894	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
Data File : VX040191.D
Acq On : 12 Feb 2024 11:57
Operator : JC/MD
Sample : P1411-01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-11929

Quant Time: Feb 12 14:22:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

