

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
 Data File : VX040197.D
 Acq On : 12 Feb 2024 14:19
 Operator : JC/MD
 Sample : P1438-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-TB-20240206

Quant Time: Feb 12 14:44:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84940	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	143918	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141040	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	74722	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56436	44.661	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.320%
35) Dibromofluoromethane	5.391	113	44110	43.889	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.780%
50) Toluene-d8	8.647	98	177267	47.151	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.300%
62) 4-Bromofluorobenzene	11.079	95	73122	47.689	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.380%
Target Compounds						
16) Acetone	2.374	43	1031	1.987	ug/l	# 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
Data File : VX040197.D
Acq On : 12 Feb 2024 14:19
Operator : JC/MD
Sample : P1438-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-TB-20240206

Quant Time: Feb 12 14:44:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

