

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
 Data File : VX040201.D
 Acq On : 12 Feb 2024 15:51
 Operator : JC/MD
 Sample : P1438-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-GW-718-720

Quant Time: Feb 12 16:17:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77324	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	131384	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66498	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	51158	44.472	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.940%
35) Dibromofluoromethane	5.385	113	41962	45.734	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.460%
50) Toluene-d8	8.647	98	160768	46.842	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.680%
62) 4-Bromofluorobenzene	11.079	95	65462	46.766	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.540%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	2489	5.268	ug/l	# 89
44) Trichloroethene	7.129	130	949	0.957	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
Data File : VX040201.D
Acq On : 12 Feb 2024 15:51
Operator : JC/MD
Sample : P1438-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-GW-718-720

Quant Time: Feb 12 16:17:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

