

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
 Data File : VX040202.D
 Acq On : 12 Feb 2024 16:14
 Operator : JC/MD
 Sample : P1438-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-GW-738-740

Quant Time: Feb 12 17:21:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77305	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	131248	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65996	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50905	44.263	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.520%
35) Dibromofluoromethane	5.385	113	40041	43.686	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.380%
50) Toluene-d8	8.647	98	156755	45.720	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.440%#
62) 4-Bromofluorobenzene	11.079	95	63718	45.568	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.140%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2353	4.982	ug/l	97
44) Trichloroethene	7.123	130	28192	28.465	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
Data File : VX040202.D
Acq On : 12 Feb 2024 16:14
Operator : JC/MD
Sample : P1438-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-GW-738-740

Quant Time: Feb 12 17:21:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

