

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
 Data File : VX040203.D
 Acq On : 12 Feb 2024 16:37
 Operator : JC/MD
 Sample : P1372-04 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CG-MW-17-ch3

Quant Time: Feb 12 17:22:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	141239	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	53893	44.486	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.980%
35) Dibromofluoromethane	5.385	113	44055	44.665	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.340%
50) Toluene-d8	8.647	98	172194	46.670	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.340%
62) 4-Bromofluorobenzene	11.079	95	68598	45.587	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.180%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	4582	9.209	ug/l #	87
40) Benzene	6.037	78	57507	14.883	ug/l	99
67) Ethyl Benzene	10.195	91	4540	0.919	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
Data File : VX040203.D
Acq On : 12 Feb 2024 16:37
Operator : JC/MD
Sample : P1372-04 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CG-MW-17-ch3

Quant Time: Feb 12 17:22:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

