

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
 Data File : VX040207.D
 Acq On : 12 Feb 2024 18:09
 Operator : JC/MD
 Sample : P1438-02DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-GW-698-700DL

Quant Time: Feb 12 18:34:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74868	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	128197	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65404	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49475	44.419	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.840%
35) Dibromofluoromethane	5.385	113	40299	45.014	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.020%
50) Toluene-d8	8.647	98	158616	47.364	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.720%
62) 4-Bromofluorobenzene	11.079	95	64024	46.876	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.760%
Target Compounds						
44) Trichloroethene	7.129	130	61650	63.728	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
Data File : VX040207.D
Acq On : 12 Feb 2024 18:09
Operator : JC/MD
Sample : P1438-02DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-GW-698-700DL

Quant Time: Feb 12 18:34:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

