

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
 Data File : VX040208.D
 Acq On : 12 Feb 2024 18:32
 Operator : JC/MD
 Sample : P1371-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T60-AOP

Quant Time: Feb 13 00:55:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	126075	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	49824	45.145	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.300%
35) Dibromofluoromethane	5.379	113	38963	44.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.500%
50) Toluene-d8	8.647	98	150249	45.620	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.240%#
62) 4-Bromofluorobenzene	11.079	95	62175	46.289	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.580%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	913	4.247	ug/l	100
16) Acetone	2.380	43	3704	8.172	ug/l	94
30) Chloroform	5.087	83	2463	1.467	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
Data File : VX040208.D
Acq On : 12 Feb 2024 18:32
Operator : JC/MD
Sample : P1371-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T60-AOP

Quant Time: Feb 13 00:55:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

