



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021224\
Data File : VX040184.D
Acq On : 12 Feb 2024 08:18
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Title : SW846 8260
Last Update : Mon Feb 12 00:21:14 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.3	6661	PASS
75	95	30	60	53.2	18405	PASS
95	95	100	100	100.0	34589	PASS
96	95	5	9	6.5	2231	PASS
173	174	0.00	2	0.7	198	PASS
174	95	50	100	79.1	27360	PASS
175	174	5	9	7.5	2040	PASS
176	174	95	101	96.4	26371	PASS
177	176	5	9	6.6	1731	PASS