

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
 Data File : VX044921.D
 Acq On : 12 Feb 2025 10:34
 Operator : JC/MD
 Sample : VX0212MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0212MBL01

Quant Time: Feb 13 00:46:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	90822	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	187975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	176520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	79122	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	73164	55.032	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.060%
35) Dibromofluoromethane	5.373	113	62053	50.774	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	8.641	98	233491	50.523	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.040%
62) 4-Bromofluorobenzene	11.079	95	84109	53.986	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
Data File : VX044921.D
Acq On : 12 Feb 2025 10:34
Operator : JC/MD
Sample : VX0212MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0212MBL01

Quant Time: Feb 13 00:46:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

