

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
 Data File : VX044922.D
 Acq On : 12 Feb 2025 10:57
 Operator : JC/MD
 Sample : VX0212WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0212WBL01

Quant Time: Feb 13 00:47:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	96997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	194457	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	176635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	76229	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76805	54.093	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.180%
35) Dibromofluoromethane	5.379	113	64443	50.972	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.940%
50) Toluene-d8	8.647	98	239094	50.011	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.020%
62) 4-Bromofluorobenzene	11.079	95	82286	51.055	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
Data File : VX044922.D
Acq On : 12 Feb 2025 10:57
Operator : JC/MD
Sample : VX0212WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0212WBL01

Quant Time: Feb 13 00:47:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

