

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
 Data File : VX044926.D
 Acq On : 12 Feb 2025 12:32
 Operator : JC/MD
 Sample : Q1347-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-EB-20250207

Quant Time: Feb 13 00:49:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	84780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	175612	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	159224	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69850	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	68685	55.345	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.700%
35) Dibromofluoromethane	5.385	113	57540	50.395	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.800%
50) Toluene-d8	8.646	98	216152	50.064	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.120%
62) 4-Bromofluorobenzene	11.079	95	76095	52.280	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.560%
Target Compounds						
16) Acetone	2.379	43	3681	7.378	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
Data File : VX044926.D
Acq On : 12 Feb 2025 12:32
Operator : JC/MD
Sample : Q1347-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-EB-20250207

Quant Time: Feb 13 00:49:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

