

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
 Data File : VX044928.D
 Acq On : 12 Feb 2025 13:17
 Operator : JC/MD
 Sample : Q1347-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-640-642

Quant Time: Feb 13 00:50:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	87862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	179992	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	164982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70282	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	71223	55.377	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.760%
35) Dibromofluoromethane	5.379	113	60220	51.459	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.920%
50) Toluene-d8	8.646	98	221052	49.953	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.900%
62) 4-Bromofluorobenzene	11.079	95	77408	51.888	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.780%
Target Compounds						
16) Acetone	2.385	43	5620	10.870	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
Data File : VX044928.D
Acq On : 12 Feb 2025 13:17
Operator : JC/MD
Sample : Q1347-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-640-642

Quant Time: Feb 13 00:50:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

