

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
 Data File : VX044929.D
 Acq On : 12 Feb 2025 13:40
 Operator : JC/MD
 Sample : Q1347-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-660-662

Quant Time: Feb 13 00:50:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	177786	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	164986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	70815	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	71019	55.507	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.020%
35) Dibromofluoromethane	5.379	113	59283	51.287	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.580%
50) Toluene-d8	8.647	98	221697	50.721	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.440%
62) 4-Bromofluorobenzene	11.079	95	78758	53.448	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.900%
Target Compounds						Qvalue
16) Acetone	2.386	43	7472	14.527	ug/l	93
17) Carbon Disulfide	2.502	76	1166	0.377	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
Data File : VX044929.D
Acq On : 12 Feb 2025 13:40
Operator : JC/MD
Sample : Q1347-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-660-662

Quant Time: Feb 13 00:50:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

