

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
 Data File : VX044930.D
 Acq On : 12 Feb 2025 14:03
 Operator : JC/MD
 Sample : Q1347-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-680-682

Quant Time: Feb 13 00:51:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	89084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	185691	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	170107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73552	56.404	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.800%
35) Dibromofluoromethane	5.379	113	61678	51.088	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.180%
50) Toluene-d8	8.647	98	226208	49.549	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	81128	52.713	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.420%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	17967	34.273	ug/l	99
17) Carbon Disulfide	2.502	76	878	0.279	ug/l #	80
25) 2-Butanone	4.574	43	7534	8.850	ug/l	96
59) 2-Hexanone	9.452	43	1949	1.422	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
Data File : VX044930.D
Acq On : 12 Feb 2025 14:03
Operator : JC/MD
Sample : Q1347-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-680-682

Quant Time: Feb 13 00:51:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

