

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
 Data File : VX044938.D
 Acq On : 12 Feb 2025 17:07
 Operator : JC/MD
 Sample : Q1355-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW2

Quant Time: Feb 13 00:52:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	85678	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	173415	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	155759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	62222	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69720	55.590	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.180%
35) Dibromofluoromethane	5.379	113	58379	51.778	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.560%
50) Toluene-d8	8.647	98	216198	50.709	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.420%
62) 4-Bromofluorobenzene	11.079	95	71446	49.708	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.420%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1848	3.665	ug/l #	76
27) cis-1,2-Dichloroethene	4.495	96	2073	1.588	ug/l	94
95) Naphthalene	13.780	128	2579	0.574	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
Data File : VX044938.D
Acq On : 12 Feb 2025 17:07
Operator : JC/MD
Sample : Q1355-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW2

Quant Time: Feb 13 00:52:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

