

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
 Data File : VX044939.D
 Acq On : 12 Feb 2025 17:30
 Operator : JC/MD
 Sample : PB166691TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB166691TB

Quant Time: Feb 13 00:52:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	87671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	182251	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	165182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69012	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	71535	55.741	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.480%
35) Dibromofluoromethane	5.379	113	60041	50.670	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.340%
50) Toluene-d8	8.647	98	221614	49.459	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.079	95	76306	50.516	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.040%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	14181	27.487	ug/l	97
18) Methyl Acetate	2.709	43	1769	1.087	ug/l #	81
43) Isopropyl Acetate	6.336	43	105213	34.264	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
Data File : VX044939.D
Acq On : 12 Feb 2025 17:30
Operator : JC/MD
Sample : PB166691TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB166691TB

Quant Time: Feb 13 00:52:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

