

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
 Data File : VX044945.D
 Acq On : 12 Feb 2025 19:48
 Operator : JC/MD
 Sample : Q1356-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WATER-FB

Quant Time: Feb 13 00:54:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	91004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	186395	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	167891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69229	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	73663	55.297	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.600%
35) Dibromofluoromethane	5.379	113	61114	50.429	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.860%
50) Toluene-d8	8.647	98	226342	49.392	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.780%
62) 4-Bromofluorobenzene	11.079	95	77602	50.231	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021225\
Data File : VX044945.D
Acq On : 12 Feb 2025 19:48
Operator : JC/MD
Sample : Q1356-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WATER-FB

Quant Time: Feb 13 00:54:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

