

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021419\
 Data File : VX007519.D
 Acq On : 13 Feb 2019 14:44
 Operator : JC/SP
 Sample : K1460-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-20190208

Quant Time: Feb 14 02:59:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	549486	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	845852	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	782099	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	375417	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	289254	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
35) Dibromofluoromethane	5.51	113	267112	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
50) Toluene-d8	8.71	98	994806	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
62) 4-Bromofluorobenzene	11.13	95	351179	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
Target Compounds						
16) Acetone	2.45	43	6508	2.436	ug/l #	88
64) Tetrachloroethene	9.34	164	29587	4.014	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021419\
Data File : VX007519.D
Acq On : 13 Feb 2019 14:44
Operator : JC/SP
Sample : K1460-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-20190208

Quant Time: Feb 14 02:59:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

