

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX021419\
 Data File : VX007527.D
 Acq On : 13 Feb 2019 18:17
 Operator : JC/SP
 Sample : K1458-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-7

Quant Time: Feb 14 10:55:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	540542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	843619	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	772840	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	371116	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	284269	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
35) Dibromofluoromethane	5.50	113	269784	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
50) Toluene-d8	8.71	98	982566	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
62) 4-Bromofluorobenzene	11.13	95	349359	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
Target Compounds						
16) Acetone	2.45	43	8159	3.104	ug/l #	88
27) cis-1,2-Dichloroethene	4.60	96	7642	1.273	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX021419\
Data File : VX007527.D
Acq On : 13 Feb 2019 18:17
Operator : JC/SP
Sample : K1458-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-7

Quant Time: Feb 14 10:55:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

