


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX021419\
Data File : VX007511.D
Acq On : 13 Feb 2019 10:13
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Title : SW846 8260
Last Update : Sat Feb 02 01:32:36 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	8617	PASS
75	95	30	60	47.9	23002	PASS
95	95	100	100	100.0	48045	PASS
96	95	5	9	6.8	3264	PASS
173	174	0.00	2	0.9	356	PASS
174	95	50	100	85.4	41042	PASS
175	174	5	9	8.0	3298	PASS
176	174	95	101	95.6	39242	PASS
177	176	5	9	7.1	2804	PASS