

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
 Data File : VX040218.D
 Acq On : 14 Feb 2024 14:24
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 15 00:02:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73307	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	124334	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45770	41.968	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.940%
35) Dibromofluoromethane	5.391	113	39652	45.667	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.340%
50) Toluene-d8	8.647	98	146123	44.989	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.980%#
62) 4-Bromofluorobenzene	11.079	95	59011	44.548	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
Data File : VX040218.D
Acq On : 14 Feb 2024 14:24
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 15 00:02:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

