

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
 Data File : VX040219.D
 Acq On : 14 Feb 2024 14:47
 Operator : JC/MD
 Sample : P1465-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-GW-778-780

Quant Time: Feb 15 00:03:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	78497	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	128393	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124548	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	49529	42.412	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.820%
35) Dibromofluoromethane	5.391	113	42496	47.395	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	8.647	98	155260	46.291	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.580%
62) 4-Bromofluorobenzene	11.079	95	61546	44.993	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.980%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	2107	4.393	ug/l	91
44) Trichloroethene	7.129	130	7359	7.595	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
Data File : VX040219.D
Acq On : 14 Feb 2024 14:47
Operator : JC/MD
Sample : P1465-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-GW-778-780

Quant Time: Feb 15 00:03:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

