

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
 Data File : VX040222.D
 Acq On : 14 Feb 2024 15:55
 Operator : JC/MD
 Sample : P1371-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T180-AOP

Quant Time: Feb 15 00:04:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74972	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	123918	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122039	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	47726	42.790	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.580%
35) Dibromofluoromethane	5.391	113	40603	46.919	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.840%
50) Toluene-d8	8.647	98	151491	46.798	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.600%
62) 4-Bromofluorobenzene	11.079	95	59865	45.345	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.680%
Target Compounds						
16) Acetone	2.373	43	1333	2.910	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
Data File : VX040222.D
Acq On : 14 Feb 2024 15:55
Operator : JC/MD
Sample : P1371-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T180-AOP

Quant Time: Feb 15 00:04:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

