

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
 Data File : VX040229.D
 Acq On : 14 Feb 2024 18:36
 Operator : JC/MD
 Sample : P1445-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Feb 15 00:06:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 12 00:21:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70826	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	120019	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117122	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46398	44.034	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.060%
35) Dibromofluoromethane	5.385	113	38809	46.303	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.600%
50) Toluene-d8	8.647	98	145461	46.395	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.800%
62) 4-Bromofluorobenzene	11.079	95	56491	44.179	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX021424\
Data File : VX040229.D
Acq On : 14 Feb 2024 18:36
Operator : JC/MD
Sample : P1445-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Feb 15 00:06:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020924W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 12 00:21:14 2024
Response via : Initial Calibration

